



Ref. Certif. No.

L8164C

Issued under the International Laboratory Accreditation Cooperation scheme for the mutual recognition of test reports and certificates issued by ISO/IEC 17025 accredited testing laboratories

Ref. Certificate No.:

L8164C

Type of product:

Laparo-Endoscopic Insufflator

Name and address of the manufacturer/applicant:

Palliare Ltd., Galway Business Park, Dangan, Galway, H91 P2DK, Ireland.

Model / Type ref.:

EVA15

Ratings and principal characteristics:

100-240 V, 50/60 Hz, 110 VA, Class I, Type BF applied part

Trademark (if any):



A sample of the product was tested and found to be in accordance with the following IEC/EN standard(s):

IEC 60601-1: 2005 + A1: 2012 (BS EN 60601-1: 2006 + A12: 2014)
Excluding clauses 11.7, 12.2, 14 & 15.1.

Additional information (if necessary):


Including national deviations for the United Kingdom/Europe, Canada & The United States of America.

As shown in the ILAC-MRA UKAS Test Report Ref. which forms part of this certificate:

L8164R

This ILAC-MRA/UKAS Test Certificate is issued by the ISO/EN 17025 accredited testing laboratory:

Cranage EMC Testing Ltd., UK Test Centre, Wallace Way,
Tern Valley Business Park, Market Drayton,
Shropshire, TF9 3AG, U.K.



Date: 30/09/2019

Signature: J. Jones